



苏州市计量测试院
江苏省洁净设备计量质量监督检验中心



检定证书

Verification Certificate

证书编号: 
Certificate No. 801414592

送检单位 Applicant	西安西电开关电气有限公司
计量器具名称 Name of Instrument	温湿度计
型号 / 规格 Type/Specification	TES1360A
出厂编号 No. of Instrument	200102755
制造单位 Manufacturer	TES
检定依据 Verification Regulation	JJG (苏) 99-2010《数字温湿度计检定规程》
检定结论 Conclusion	合格



批准 Approved by		胡涵星	主管
核验 Checked by		王恩淳	
检定 Verified by		周鑫锋	

检定日期 Date of Verification	2020	年 Year	06	月 Month	01	日 Day
有效期至 Valid to	2021	年 Year	05	月 Month	31	日 Day

计量检定机构授权证书号: (苏) 法计(2017)2055号
Authorization certificate No. (2017)2055.

地址: 苏州市吴中区文曲路69号
Address

网址: www.szjl.com.cn
HTTP

电话: 0512-65230843
Tel.

邮编: 215128
Post Code

传真: 0512-65230844
Fax.

电子邮件: ywgl@szjl.com.cn
E-mail



证书编号：
Certificate No.



801414592

本次检定所使用的社会公用计量标准：
Measurement standards used in the verification

计量标准名称 Name	测量范围 Measurement Range	不确定度/准确度 Uncertainty/Accuracy	证书编号/有效期 Certificate No./Valid Date
精密露点仪标准装置	温度：(5~60)℃，湿度：(30~95)%RH 温度：(5~60)℃，湿度：(30~95)%RH	温度：±0.1℃ 湿度：(0.9~1.5)%RH k=2 温度：±0.1℃ 湿度：(0.9~1.5)%RH k=2	(2014) 江苏省社公量标 证字第290号/ 2023-05-27

本次检定所使用的主要计量器具：
Main measurement standards used in the verification

名称 Name	编号 No.	测量范围 Measure Range	不确定度或准确度等级 或最大允许误差 Uncertainty or Accuracy Class or MPE.	溯源机构 Traceability Mechanism	证书编号/ 有效期 Certificate No./Due Date
精密露点仪 Precision Dew point Meter	33126	(-20~40)℃DP/(-10~60)℃	MPE: ±0.2℃DP/±0.1℃ MPE: ±0.2℃DP/±0.1℃	苏州市计量 测试院	801253535/ 2020-08-15

检定地点和环境条件：

Location and environmental condition for the verification

地点：苏州市文曲路69号 (No. 69 Wenqu road, Suzhou) 本院实验室210

Location

温度：(20 ~ 22)℃

Ambient Temperature

湿度：(57 % ~ 60 %)RH

Humidity

其他： /

Others

注：

Statement

- 1、未经本所批准授权，不得部分采用本证书。Partly using this certificate will not be admitted unless allowed by SIM.
- 2、本证书的检定结果仅对所检定器具有效。The results of this certificate are only responsible for the item Verified.
- 3、本证书未加盖检定专用章无效。The certificate is invalid without official stamp.
- 4、证书结果中"P"代表"合格"，"F"代表"不合格"，"N/A"代表"不适用"。In the results of the calibration, "P" stands for "Pass", "F" stands for "Fail", and "N/A" stands for "No applicable".

检定证书续页专用

Continue page of verification certificate

第 2 页 共 3 页

Page of



证书编号:
Certificate No.



801414592

检定结果/说明:

Results of verification and additional explanation

1. 外观和功能检查(Appearance and Function Check)

序号 No.	测量结果 Results	技术要求 Technical Requirements	结论 Pass/Fail
1	符合要求	外形结构完好。数字指示面板不应有影响读数的缺陷，数字显示不应出现间隔跳动的现象。	P

2. 计量特性测量结果(The Results Unrelated to Metrological Specification)

单位 (Unit) : °C

测量点 Measuring Points	被检示值 Sample Values	示值误差 Error	允许误差 MPE	测量不确定度 Uncertainty in Measurement $U, k=2$	结论 Pass/Fail
10.0	10.1	+0.1	± 2.0	0.4	P
20.0	20.2	+0.2	± 2.0	0.4	P
30.0	30.2	+0.2	± 2.0	0.4	P

单位 (Unit) : %RH

测量点 Measuring Points	被检示值 Sample Values	示值误差 Error	允许误差 MPE	测量不确定度 Uncertainty in Measurement $U, k=2$	结论 Pass/Fail
80.0	82.3	+2.3	± 7.0	1.5	P
60.0	60.3	+0.3	± 5.0	1.3	P
40.0	39.5	-0.5	± 5.0	1.2	P

以下空白
Blank Below

检定人员:
Verified by

周鑫锋

周鑫锋

核验人员:
Checked by

王恩淳

王恩淳

检定证书续页专用

Continue page of verification certificate

第 3 页 共 3 页
Page of